

	Certificate of Analysis	COA No: CA_XBE-0002-4
		Version: 07

Low DNA Taq HS 5 U/μL Suitable for Research and further Manufacturing Use	Catalog No:	MDX009
	Lot No:	EN013-B356550
	Shipping / Storage Conditions:	-20°C
	Component Lot No:	IM-425110A
	Expiry date:	November 2027

Quality Control Parameters

Heat-activated, thermostable DNA polymerase suited to amplification of bacterial and fungal DNA

Analysis	Specification	Result
Activity	Quantitative PCR analysis amplifying 1 gene from a dilution series of enzyme under standard conditions. Cq and melting profiles must be consistent for the test and reference sample with ± 0.5 Cq variance.	Passed
Sensitivity	Quantitative PCR analysis amplifying 1 gene from a dilution series of mouse cDNA under standard conditions. Cq and melting profiles must be consistent for the test and reference sample with ± 0.5 Cq variance. A 3Kb fragment is amplified with a dilution series of Lambda DNA, using standard conditions and 30 cycles. Single distinct bands were observed with agarose gel electrophoresis (ethidium stained).	Passed
Heat activation	A 125bp fragment is amplified with a dilution series of enzyme, using 4 heat activation times and 30 cycles. Single distinct bands were observed, at the appropriate activation time, with agarose gel electrophoresis (ethidium stained).	Passed
Purity	Densitometric analysis of SDS-Page. Purity must be higher than 90%	98.9 %
DNA contamination	Quantitative PCR analysis with no template. Presence of <i>E. coli</i> and mouse genomic DNA checked. Test sample must amplify in line with a reference sample.	Passed
DNase contamination	Incubation of a 1Kb double stranded DNA fragment. Incubation for 4 hours at 37°C with dilution series of DNase I. Analysed by agarose gel electrophoresis. Test sample must show less degradation than the limit of detection 2.5×10^{-3} U DNase.	Passed

QA / QC Representative: *Xinghan*

X.Chen

Date: 28th October 2025

United Kingdom

Tel: +44 (0)20 8830 5300
Fax: +44 (0)20 8452 2822

USA

Tel: +1 901.382.8716
Fax: +1 901.382.0027

Germany

Tel: +49 (0)3371 60222 00
Fax: +49 (0)3371 60222 01

	Certificate of Analysis	COA No: CA XBB-0002-2
		Version: 05

Low DNA Reaction Buffer, 10x For research or further manufacturing use only	Catalog No:	MDX009
	Lot No:	EN013-B356550
	Storage Conditions:	-20°C
	Component Lot No:	IB-525210A
	Expiry date:	November 2027

Quality Control Parameters

Low DNA Reaction Buffer 10x is a combination of the latest advances in buffer chemistry together with enhancers and stabilizers at optimal concentrations. It has been designed for use with Low DNA Taq HS making it ideal for PCR of low-copy bacterial targets to avoid false-positive amplification, such as in water testing

Analysis	Specification	Result
Functional	Fragment of size 800bp was amplified with a dilution series of Low DNA Taq, using standard conditions and 30 cycles. Single distinct bands were observed with agarose gel electrophoresis (ethidium stained).	Passed
DNA contamination	Quantitative PCR analysis with no template. Presence of <i>E. coli</i> and mouse genomic DNA checked. Test sample must amplify in line with a reference sample.	Passed
DNase contamination	Incubation of a 1Kb double stranded DNA fragment. Incubation for 4 hours at 37°C with dilution series of DNase I. Analysed by agarose gel electrophoresis. Test sample must show less degradation than the limit of detection 2.5×10^{-3} U DNase.	Passed

QA / QC Representative: 

X.Chen

Date: 28th October 2025

United Kingdom

Tel: +44 (0)20 8830 5300
Fax: +44 (0)20 8452 2822

USA

Tel: +1 901.382.8716
Fax: +1 901.382.0027

Germany

Tel: +49 (0)3371 60222 00
Fax: +49 (0)3371 60222 01

	Certificate of Analysis	COA No: CA_XBB-0014
		Version: 09

MgCl₂ Solution, 50mM For research or further manufacturing use only	Catalog No:	MDX009
	Lot No:	EN013-B356550
	Storage Conditions:	-20°C
	Component Lot No:	MG-525110D
	Expiry date:	November 2027

Quality Control Parameters

Analysis	Specification	Result
Functional	Fragments of sizes 800bp and 3000bp are amplified with a dilution series of BIOTAQ™ DNA Polymerase, using standard conditions and 30 cycles. Single distinct bands were observed with agarose gel electrophoresis (ethidium stained).	Passed
DNA contamination	Quantitative PCR analysis with no template. Presence of <i>E. coli</i> and mouse genomic DNA checked. Test sample must amplify in line with a reference sample.	Passed
DNase contamination	Incubation of a 1Kb double stranded DNA fragment. Incubation for 4 hours at 37°C with dilution series of DNase I. Analysed by agarose gel electrophoresis. Test sample must show less degradation than the limit of detection 2.5×10^{-3} U DNase.	Passed

QA / QC Representative: 

X.Chen

Date: 28th October 2025

United Kingdom

Tel: +44 (0)20 8830 5300
Fax: +44 (0)20 8452 2822

USA

Tel: +1 901.382.8716
Fax: +1 901.382.0027

Germany

Tel: +49 (0)3371 60222 00
Fax: +49 (0)3371 60222 01